



# DECLARATION OF CONFORMITY TO COUNCIL DIRECTIVE 93/42/EEC CONCERNING MEDICAL DEVICES

## Appendix: list of (harmonised - EN) standards

NO.	Reference	Title
1	IEC60601-1:2012	Medical electrical equipment - Part 1: General requirements for basic safety and essential performance
2	IEC60601-1-2:2014	Medical electrical equipment - Part 1-2: General requirements for basic safety and essential performance - Collateral Standard: Electromagnetic disturbances - Requirements and tests
3	IEC 60601-1-6:2013	Medical electrical equipment-Part 1-6:General requirements for basic safety and essential performance-Collateral Standard: Usability
4	IEC 60601-1-11:2015	Medical electrical equipment – Part 1-11: General requirements for basic safety and essential performance - Collateral standard: Requirements for medical electrical equipment and medical electrical systems used in the home healthcare environment
5	IEC 62366-1:2015	Medical devices - Application of usability engineering to medical devices
6	ISO 80601-2-61: 2017	Medical electrical equipment —Part 2-61: Particular requirements for basic safety and essential performance of pulse oximeter equipment
7	IEC 62304:2015	Medical device software-Software life-cycle processes



Benannt durch/Designated by  
 Zentralstelle der Länder  
 für Gesundheitsschutz  
 bei Arzneimitteln und  
 Medizinprodukten  
 www.zlg.de  
 ZLG-BS-244.10.08



Product Service

# EC Certificate

Full Quality Assurance System

Directive 93/42/EEC on Medical Devices (MDD), Annex II excluding (4)  
 (Devices in Class IIa, IIb or III)

**No. G1 050972 0050 Rev. 04**

**Manufacturer:**

**Contec Medical Systems Co., Ltd.**

No.112 Qinhuang West Street  
 Economic& Technical Development Zone  
 066004 Qinhuangdao, Hebei Province  
 PEOPLE'S REPUBLIC OF CHINA

**Product Category(ies):**

**Patient Monitor, Fetal Monitor, B-Ultrasound Diagnostic System, Pulse Oximeter, Electrocardiograph, Pocket Fetal Doppler, Visual Electronic Stethoscope, Multi-functional Visual Stethoscope, Dynamic ECG Systems, Digital Brain Electric Activity Mapping, Infusion Pump, Spirometer, Ambulatory Blood Pressure Monitor, Electronic Sphygmomanometer, EMG/EP System, Portable ECG Monitor, Temperature Probe, Pulse Oximeter Probe, Tele Pulse Oximeter, Tele Breather, Multi-parameter Vital Signs Monitor, Sleep apnea screen meter, Oxygen concentrator, ECG Workstation, Wearable Monitor, Mesh Nebulizer, Capnograph and Infrared Thermometer.**

The Certification Body of TÜV SÜD Product Service GmbH declares that the aforementioned manufacturer has implemented a quality assurance system for design, manufacture and final inspection of the respective devices / device categories in accordance with MDD Annex II. This quality assurance system conforms to the requirements of this Directive and is subject to periodical surveillance. For marketing of class III devices an additional Annex II (4) certificate is mandatory. See also notes overleaf.

**Report No.:**

BJ20090203

**Valid from:**

2020-06-17

**Valid until:**

2024-05-26

**Date,**

2020-06-17

Christoph Dicks  
 Head of Certification/Notified Body