



LUX-SUTURES S.A. 22, Gruuss-Strooss 9991 Weiswampach LUXEMBOURG info@luxsutures.com Tel.: 00352-2030-1449

01 March 2021

## **DECLARATION OF CONFORMITY**

## Manufacturer: LUX-SUTURES S.A. 22, Gruuss-Strooss, 9991 Weiswampach LUXEMBOURG

We herewith declare under our sole responsibility that the product:

## LUXAMID

## Polyamide - Nylon - sterile non-absorbable suture in foil packs In atraumatic needle and without needle GMDN code: 38000

Needle Description		
Characteristics	Description	
Needle Types & Codes	Taper point	1/2 circle (HR), 3/8 circle (DR), 5/8 circle (FR), J needle (JR), Straight (GR)
	Reverse cutting	1/2 circle (HS), 3/8 circle (DS), Straight (GS)
	Reverse cutting Easy Pass	3/8 circle (DS)
	Tapercut	1/2 circle (HRT), 3/8 circle (DRT), J needle (JRT)
	Cutting	3/8 circle (NS), 1/2 circle (NH)
	Spatula	1/2, 3/8, 1/4 circle or straight
	Blunt point	1/2 circle (HRB), 3/8 circle (DRB)
	Micro needle – taper point	3/8 circle
	Without needle	n/a
Number of Needles	1 or 2	
Needle Angle Range	1/4, 3/8, 5/8, 1/2, J, Straight	
Needle Lenght Range	3 mm to 100 mm	
Suture Thread Characteristics		
Suture Color	Black or blue	
Suture Gauze sizes (USP)	2, 1, 0, 2/0, 3/0, 4/0, 5/0, 6/0, 7/0, 8/0, 9/0, 10/0, 11/0	
Suture Lenght Range	10 cm - 250 cm	
Quanitity per Pouch	1-17	
Presentation	Sterile sutures in foil packs	

LUX-DOC-C-NY

2021-03-01

VAT No.: LU18518363

IBAN: BE49 6343 2171 0171 BIC (SWIFT): BNAGBEBB





LUX-SUTURES S.A. 22, Gruuss-Strooss 9991 Weiswampach LUXEMBOURG info@luxsutures.com Tel.: 00352-2030-1449

is a medical device and has been classified as Class IIb (Annex IX Rule 8). It conforms to the essential requirements and provisions of the Council Directive 93/42/EEC as amended by Directive 2007/47/EC.

It is subject to the conformity assessment procedure set out in Annex II of Directive 93/42/EEC as amended by Directive 2007/47/EC under the supervision of the National Evaluation Center of Quality and Technology in Health, EKAPTY, Notified body, with identification number 0653.

Name and function of the responsible person:

Nico WEBER QA & RA Manager

tober



LUX-DOC-C-NY

2021-03-01