

B&E KOREA CO., LTD.

#802, #807, B-dong, 60, Haan-ro, Gwangmyeong-si, Gyeonggi-do, Republic of Korea, 14322 TEL 82-2-2083-2820 FAX. 82-6112-2816

CERTIFICATE OF ANALYSIS

B&E Etch-37

	Lot Number BEE23009	Pack 100	Production Date 2023.08.15	Expiry Date 2026.08.14
Item	Method	Criteria		Result
1. Appearance	By bare eye	No break, spot, split, defect, crack, hole, extraneous substances at the package		Pass
2. PH	By ph measuring method	0.5~1.5		0.76
3. Leakage	By bare eye	No leakage		Pass

1. COMPOSITION

Water 30-70%

Phosphoric acid 10-40% ■ Sugar classification 1-10%

2. QUALITY

- 1) Contains bio-compatible polymeric materials
- 2) Completely residue-free after spray with water
- 3) Can be placed exactly where you want it, because of excellent viscosity
- 3. APPEARANCE: Gel type

4. FLASH POINT AND METHOD OF USE

B&E Etch-37 is a pre-treatment agent which helps to bonding agent having a strong bonding strength between and teeth.

5. SPECIFICATIONS

a) Class Ⅱ a

according to the Rule 6 of Classification Criteria, Annex IX, MDD 93/42/EEC

b) Presented as FineEtch

5ml / syringe, 1.2ml / syringe

6. LIFE OF PRODUCTS: 3 years from the manufactured date

7. HANDLING: This material is intended for use by properly trained dental professionals only

8. STORAGE : Store in shady and cool place. (5°C~24°C)

BAEK KYU HYUN / President