



B&E KOREA CO., LTD.

#802, #807, B-dong, 60, Haan-ro, Gwangmyeong-si, Gyeonggi-do, Republic of Korea, 14322
TEL 82-2-2083-2820 FAX. 82-6112-2816

CERTIFICATE OF ANALYSIS

B&E Block Dam

Lot Number	Pack	Production Date	Expiry Date
BDB23008	10	2023.08.15	2025.08.14

1. COMPOSITION

■ Urethane dimethacrylate	60-80%
■ Silicone dioxide	10-30%
■ Pigment	1-10%
■ Other ingredients determined	to 100

2. QUALITY

- 1) Light Cured resin barrier for isolating tissue adjacent to teeth being whitened.
- 2) For Single tooth bleaching, B&E Block-Dam may be used protect closed teeth

3. APPEARANCE : syringe

4. INDICATION FOR USE

Remove cap from syringe, Place tip onto syringe.
Turn tip 90degree to look in place
Apply the block Dam, light curing for 20sec

5. SPECIFICATIONS

- a) Class II a
according to the Rule 5 of Classification Criteria, Annex IX, MDD 93/42/EEC

6. LIFE OF PRODUCTS : 2 years from the manufactured date

7. HANDLING : This material is intended for use by properly trained dental professionals only

8. STORAGE : Store in dry and cool place. (2°C~27°C)

BAEK KYU HYUN / President