



B&E KOREA CO., LTD.

#802, #807, B-dong, 60, Haan-ro, Gwangmyeong-si, Gyeonggi-do, Republic of Korea, 14322
TEL 82-2-2083-2820 FAX. 82-6112-2816

CERTIFICATE OF ANALYSIS

X-Temp

Product Name	Lot Number	Pack	Production Date	Expiry Date
X-Temp Yellow	XTY23007	10	2023.08.15	2025.08.14

Item	Method	Criteria	Result
1. Appearance	By bare eye	No extraneous substances at the device	Pass
2. Packaging	By bare eye	No break, spot, split, defect, crack, hole, extraneous substances at the package	Pass
3. Capacity	By capacity measuring method	± 0.5 %	3.0g
4. Depth of curing	By ISO4049	Over 1.5mm	2.7 mm
5. Setting time	By ISO4049	less than 20 sec	10 sec

1. COMPOSITION

- Urthane dimethacrylate 20-70%
- Silicone dioxide 10-30%
- Pigment 1-10%

2. QUALITY

- 1) Simple to apply : None sticky placement
- 2) Easy to remove in one piece
- 3) Suitable for inlays and onlays
- 4) Durable and tight marginal seal

3. APPEARANCE : Paste (Tooth color, Blu color)

4. FLASH POINT AND METHOD OF USE

X-Temp is light-cured elastic materials for temporary restorations.

5. SPECIFICATIONS

a) Class II a

according to the Rule 7 of Classification Criteria, Annex IX, MDD 93/42/EEC

b) Shade : Tooth color, Blue color

6. LIFE OF PRODUCTS : 2 years from the manufactured date

7. HANDLING : This material is intended for use by properly trained dental professionals only

8. STORAGE : Store in dry and cool place. (2°C~27°C)



BAEK KYU HYUN/ President